

Intertek ETL SEMKO

7250 Hudson Blvd., suite 100, oakdale MN 55128

EMC VERIFICATION No. 3070302.011 - 172

EQUIPMENT UNDER TEST

Type of equipment	Portable Turbidity, Color, and Chlorine Meters (Tri-Meters)
Brand name	LaMotte
Type/Model	TC3000e and TC3000i
Manufacturer	Apprise Technologies Inc.
Tested by request of	LaMotte Company

STANDARDS

TEST REPORT No. 3070302.011

Emissions

EN61326:1997, Class B
EN61000-3-2:2000
EN61000-3-3:1995

Immunity

EN 61326:1997, Minimum Immunity Requirements

SUMMARY OF RESULTS

We confirm that the product tested and our review of the above numbered report without reasonable doubt will fulfill the requirements concerning electromagnetic compatibility according to the above-mentioned standards harmonized with the EMC Directive 89/336/EEC.

EMC Department

Date of issue: January 31, 2005

Signature: _____

Yuriy Litvinov, Site Manager

